
H-3810


H-3812


H-3813B


H-3814


H-3811


H-3056.3F


H-3056.1




H-3056.4
H-3056.2


H-3817


H-3816.1M


H-3817.20


H-3056.11



H-3056.10

Blaine Air Permeability Apparatus
ASTM C204; AASHTO T153

The Blaine Air Permeability Apparatus determines the fineness of Portland cement in terms of specific surface area expressed as the total surface area in square centimeters per gram of mortar or cement. The Blaine works by drawing a definite quantity of air through of bed of cement exhibiting a definitive porosity value. The number and size of the pores in a prepared bed of definite porosity is a function of the size of the particles and determines the rate of airflow through the bed. This procedure is outlined in ASTM C204, Method A and AASHTO T153. The Humboldt Blaine Air Permeability Apparatus consists of: calibrated U-tube manometer, ground glass joint, stainless steel test cell and plunger, rubber aspirator bulb and perforated disc. Includes an 8 oz. (226.8g) bottle of red manometer fluid, filter paper, wood block for holding test cell during filling and a funnel. Mounted on finished wood panel with rubber-footed base. To perform this test, NIST Portland Cement #114R is required by the ASTM standard for calibration. See Accessories H-3817 and H-3817.20.

Blaine Air Permeability Apparatus **H-3810**

Ship wt. 7lbs. (3.1kg)

H-3810 Accessories & Replacement Parts

Description	Model
Rubber bulb	H-3811
Cell and plunger	H-3812
Cell and plunger, Calibrated	H-3812CAL
Perforated brass disc	H-3813B
Perforated stainless steel disc	H-3813S
Manometer fluid, 8oz (240ml)	H-3814
Manometer u-tube, calibrated	H-3815
Filter paper discs, medium retentive, 1.27cm, pkg 1000	H-3816.1M

Portland Cement Fineness Standard, SRM 114R

This Standard Reference Material (SRM) is used in calibrating fineness testing equipment according to ASTM Standard Methods. The SRM unit consists of a glass vial with plastic caps containing powdered cement (each vial is contained in a sealed foil bag). Each vial contains approximately 5g of cement.

SRM 114R 1 vial **H-3817**
SRM 114R, package of 20 **H-3817.20**

Ship wt. 0.5lbs. (0.2kg)

Blaine Apparatus, Semi-Automatic
ASTM C204; AASHTO T153, EN 196; BS 4550

Determines fineness of cement in terms of specific surface expressed as total surface area square centimeters per gram. To obtain the most accurate results, the test should be performed in a temperature-controlled environment. The Semi-automatic Blaine Apparatus provides more accuracy and precision than provided by the manual Blaine Apparatus. The device uses an automatic pump and timer to evaluate the time precisely. Calibration of this unit is done by using various reference sands. Unit includes: the unit with an electric pump and timer; measuring cell, filter papers (12.8mm, 1000pk.; fill oil (50ml); plug; thermometer; brush and funnel.

Blaine Apparatus, Semi-Automatic 120/220V 50/60 Hz **H-3056.3F**

Ship wt. 15lbs. (6.8kg)

H-3056.3F Accessories & Replacement Parts

Description	Model
Calibration Sand, Coarse, 125g	H-3056.2
Calibration Sand, Fine, 125g	H-3056.4
Fill Oil, 50ml	H-3056.5
Light Grease	H-3056.10
U-Shaped Tube for H-3056.3F	H-3056.6
Filter Papers, 12.8mm (1000)	H-3056.1
Measuring Cell and Plunger, 1.8cm^3	H-3056.11
Perforated Disc	H-3056.8
Tamper	H-3056.13