

Rotor oscilante 48 tubos de recolección de sangre de 5/7 ml para centrifuga de mesa refrigerada de alta velocidad

Código: B2-01-010384-R22

Marca: N/D

[Ver producto en Cotecno](#) | [Solicitar cotización](#)

[Detalle](#) | [Especificaciones](#) | [Aplicaciones](#)

Detalle

Swing Rotor

Especificaciones técnicas

Especificación	-R22
Tipo de rotor	Rotor oscilante
Volumen	48*5/7ml tubos de recolección de sangre
Velocidad máxima (rpm)	4000
Fuerza centrífuga relativa máxima (xg)	2952
Adaptadores	tubo de recolección de sangre de 2/3/4ml

Aplicaciones

1. Confirm the tubes you will spin

Compare the rotor's number of positions and the blood collection tubes it accepts with the samples you process in each run.

- **Volume:** 48*5/7ml blood collection tube

If you work with tubes of another size, check the adapters the rotor accepts.

- **Adapters:** 2/3/4ml blood collection tube

2. Check speed and centrifugal force

Check that the rotor's maximum speed and maximum relative centrifugal force meet what your separation protocol requires.

- **Max. Speed(rpm):** 4000
- **Max. RCF(xg):** 2952

3. Choose the rotor type

Choose between an angle rotor and a swing rotor according to the separation you need, and confirm that the rotor matches your centrifuge.

- **Rotor Type:** Swing Rotor

Siguiente paso: [ver producto en Cotecno](#) | [solicitar cotización](#)